

## United States Patent and Trademark Office



UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER OF PATENTS AND TRADEMARKS Washington, D.C. 20231 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
09/440,106	11/15/1999	CHARLES S. TAYLOR	GUID-003DIV2	2546	
	11/15/1999 CHARLES S. TAYLOR  2590 02/27/2002  ehlke FIELD & FRANCES LLP d Road Suite 200		·····		
Kimberly Zuehlke BOZICEVIC FIELD & FRANCES LLP 200 Middlefield Road Suite 200			EXAMINER		
			NASSER, ROBERT L		
Menlo Park, CA 94025			ART UNIT	PAPER NUMBER	
			3736		

DATE MAILED: 02/27/2002

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No. 09/440,106 Applicant(s)

Taylor et al



Office Action Cummons			
Office Action Summary	Examiner Robert Nasser	Art Unit 3736	
The MAILING DATE of this communication appears	on the cover sheet with the corre	spondence addre	·ss
Period for Reply  A SHORTENED STATUTORY PERIOD FOR REPLY IS SET  THE MAILING DATE OF THIS COMMUNICATION.			
<ul> <li>Extensions of time may be available under the provisions of 37 C after SIX (6) MONTHS from the mailing date of this community.</li> <li>If the period for reply specified above is less than thirty (30) day be considered timely.</li> <li>If NO period for reply is specified above, the maximum statutory communication.</li> <li>Failure to reply within the set or extended period for reply will, be any reply received by the Office later than three months after the earned patent term adjustment. See 37 CFR 1.704(b).</li> </ul>	period will apply and will expire SIX (	m of thirty (30) da (6) MONTHS from	ays will the mailing date of this D (35 U.S.C. § 133).
Status	2001		
1) X Responsive to communication(s) filed on <u>Aug 29,</u>			
Za/ This detict is the table	ction is non-final.		ito io
3) Since this application is in condition for allowance closed in accordance with the practice under Ex p	except for formal matters, pros parte Quayle, 1935 C.D. 11; 453	ecution as to th 3 O.G. 213.	e merits is
Disposition of Claims	ic/a	re nending in th	e application.
4) 💢 Claim(s) <u>1-22</u>		that are a	ideration
4a) Of the above, claim(s)	IS/8	are withdrawn t	rom consideration.
5) Claim(s)		_ is/are allowed	
6) 💢 Claim(s) <u>1-22</u>		_ is/are rejected	<b>J</b> .
7) Claim(s)		_ is/are objecte	d to.
8) Claims	are subject to rest	riction and/or el	ection requirement.
Application Papers			
9) The specification is objected to by the Examiner.			
10) The drawing(s) filed on is/a			
11) The proposed drawing correction filed on		d b)□ disappro	oved.
12) The oath or declaration is objected to by the Exa			
Priority under 35 U.S.C. § 119  13)		a)-(d).	
1. Certified copies of the priority documents h	nave been received.		
2. Certified copies of the priority documents h	nave been received in Application	No	•
3. Copies of the certified copies of the priority	/ documents have been received ureau (PCT Rule 17.2(a)).	in this National	Stage
*See the attached detailed Office action for a list of 14) Acknowledgement is made of a claim for domes			
14) Acknowledgement is made of a claim for domes	ide priority under do o.o.o. 3 1 1	= 1=1:	
Attachment(s)			
15) X Notice of References Cited (PTO-892)	18) X Interview Summary (PTO-413) Pa		
16) Notice of Draftsperson's Patent Drawing Review (PTO-948)	19) Notice of Informal Patent Applicat	tion (PTO-152)	
17) Information Disclosure Statement(s) (PTO-1449) Paper No(s).	20) Other:		

Application/Control Number: 09/440106 Page 2

Art Unit: 3736

The examiner notes that the preliminary amendment requested that claims 25-32 be canceled and claims 33-40 be entered. However, the only claims previously of record in this application were claims 1-14. Accordingly, claims 1-14 are still pending and claims 33-40 have been renumbered as claims 15-22. Applicant should confirm this in response to this action.

Claims 15-22 are rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-7 of U.S. Patent No. 5727969 in view of Wright et al. The previous claims do not recite attaching the suction housing to a stable object. This is clearly taught by Wright et al. Therefore, it would have been obvious to modify the previous invention to use such a fixation system, to better ensure heart stability.

Claim 16 is rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claims 1-7 of U.S. Patent No. 5727969 in view of the Roux et al article entitled "New Helper Instrument in Cardiac Surgery." The previous claims do not recite attaching the suction housing to a stable object. Roux et al shows a device to stabilize the heart which attaches the stabilizer to a retractor to hold the heart still. Therefore, it would have bene obvious to modify the previous invention to use such a stabilizing step, so as to ensure heart immobilization.

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless --

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who

Application/Control Number: 09/440106

Art Unit: 3736

has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371© of this title before the invention thereof by the applicant for patent.

Claims 1-22 are rejected under 35 U.S.C. 102(e) as being clearly anticipated by Wright et al, noting that the housing is dome shaped.

Claim 1-14 and 16 is rejected under 35 U.S.C. 102(e) as being clearly anticipated by Borst et al.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 15 and 17-22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Borst et al in view of Wright et al. Wright shows a dome shaped pressure applicator. It would have been obvious to modify Borst to replace the paddle with the dome shaped paddle, as it is merely the substitution of one known equivalent for another.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Robert L. Nasser Jr. whose telephone number is (703) 308-3251. The examiner can normally be reached on Monday-Thursday and alternate Fridays from 8:30 to 6:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Cary O'Connor, can be reached on (703) 308-2701. The fax phone number for this Group is (703) 308-0758.

Communications via Internet e-mail regarding this application, other than those under 35 U.S.C. 132 or which otherwise require a signature, may be used by the applicant and should be addressed to [cary.o'connor@uspto.gov].

All Internet e-mail communications will be made of record in the application file. PTO employees do not engage in Internet communications where there exists a possibility that sensitive

Application/Control Number: 09/440106 Page 4

Art Unit: 3736

information could be identified or exchanged unless the record includes a properly signed express waiver of the confidentiality requirements of 35 U.S.C. 122. This is more clearly set forth in the Interim Internet Usage Policy published in the Official Gazette of the Patent and Trademark on February 25, 1997 at 1195 OG 89.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the Group receptionist whose telephone number is (703) 308-0858.

RLN February 1, 2002

What & Masser PRIMARY EXAMINER